

*Los Angeles County Sheriff's Department
Blood Alcohol Testing Section*

DataMaster cdm report code explanations

Code in Report (Instrument usage Log or Subject Data Report)	Message on print-out from the instrument under -BREATH ANALYSIS-
I	INSTRUMENT TIME OUT
V	INVALID SAMPLE WAIT 15 MINUTES
X	INTERFERENCE DETECTED